

Multifocal intraocular lenses in cataract surgery: Literature review of benefits and side effects

Niels E. de Vries, MD, PhD, FEBO, Rudy M.M.A. Nuijts, MD, PhD

This literature review looks at the current status of multifocal intraocular lenses (IOLs) in cataract surgery. The results of implantation of multifocal IOLs of diffractive, refractive, and hybrid diffractive–refractive design are described with regard to uncorrected near and distance visual acuity and spectacle independence. The occurrence of photic phenomena and contrast sensitivity loss with multifocal IOLs are also addressed.

Financial Disclosure: Neither author has a financial or proprietary interest in any material or method mentioned.

J Cataract Refract Surg 2013; 39:268–278 © 2013 ASCRS and ESCRS

Accommodation is the ability of the eye to dynamically change its optical power to create a sharp image of distant, intermediate, and near objects on the retina.¹ Helmholtz² pioneered the theory that accommodation is the result of changes in the optical power of the crystalline lens as a result of changes in the lens shape and position due to changes in tension exerted on the zonular fibers after relaxation or contraction of the ciliary muscle. As people age, the ability to accommodate decreases, resulting in presbyopia.³ This is thought to be the result of changes in the elasticity of the crystalline lens^{4,5} and in the contractility of the ciliary muscle.^{6,7} Thus, even emmetropic subjects who were spectacle independent when they were younger will become dependent on spectacles for near vision once they become presbyopic.

Apart from the age-related changes in the crystalline lens that lead to presbyopia, age-related changes in the proteins in the crystalline lens lead to cataract formation.⁸ Cataract surgery with implantation of an intraocular lens (IOL) has the potential not only to increase visual acuity, but also to change the patient's refractive state. Ideally, an IOL would allow the presbyopic patient to regain his or her ability to

accommodate. Although refilling the capsular bag with a clear but elastic substance would theoretically lead to the desirable result, experiments in this area have been unsuccessful.⁹ Similarly, a change in position of the IOL or parts of it within the optical system would change the optical power of the optical system as a whole, thus providing the patient with the ability to accommodate.¹⁰ Ultrasound studies have shown changes in the position of accommodating IOLs within the optical system in response to physiological or pharmaceutical stimuli,¹¹ although other studies have not found significant movement of these IOLs.^{12,13} In clinical practice, movement of accommodating IOLs has been shown to be insufficient to result in large changes in the power of the optical system.^{10,14}

Apart from strategies to provide IOLs with a dynamic optical power or position within the optical system, IOLs can be designed to provide 2 or more fixed optical powers. So-called multifocal IOLs have been designed to result in 2 or more coexisting retinal images in which only the image corresponding to the distance or near focal point is sharp. This concept is known as *simultaneous vision*,¹⁵ although *simultaneous imaging* would be a more appropriate term. Multifocal IOLs have 2 or more fixed adapting focal points rather than 1 and are therefore pseudoaccommodative rather than truly accommodative.

The earliest multifocal IOLs were introduced in the late 1980s.^{16,17} As presented in Table 1, multifocal IOLs using refractive, diffractive, and combinations of both optical principles have been developed. Refraction is based on a change in direction of the light ray due to a change in the optical density of the material transmitting the light ray. Diffraction is based on the observation that light that encounters a discontinuity

Submitted: February 19, 2012.

Final revision submitted: June 18, 2012.

Accepted: June 19, 2012.

From the Department of Ophthalmology, University Hospital Maastricht, the Netherlands.

Corresponding author: Niels E. de Vries, MD, PhD, FEBO, University Hospital of Maastricht, PO Box 5800, 6202 AZ Maastricht, the Netherlands. E-mail: ndv@soog.azm.nl or nielserikdevries@hotmail.com.

Table 1. Characteristics of multifocal IOLs.

IOL	Company	Design	Pupil	Near Add (D)	Toric	Aspheric
AcriLISA (366D, 376D, 536 D) AT LISA (801, 802, 809M)	Carl Zeiss Meditec	Diff + Ref	Independent	+3.75	No	Yes
Acri.LISA toric (466TD) AT LISA toric (909M)	Carl Zeiss Meditec	Diff + Ref	Independent	+3.75	Yes	Yes
Acri.Twin (733 + 737)	Acri.Tech/Carl Zeiss Meditec	Diff	Independent	+4.0	No	Yes
AcriviaReviol (BB MF 613, BB MFM 611)	VSY Biotechnology	Diff	Independent	+3.75	No	Yes
Array (SA40N, SA40NB)	Abbott Medical Optics	Ref	Dependent	+3.50	No	No
CeeOn 811E*	Pharmacia	Diff	Independent	+4.0	No	No
FineVision	Physiol	Diff, trifocal	Dependent	+1.75, +3.50	No	Yes
LentisMplus (LS-312MF 15)	Oculentis GmbH	Ref, sector-shaped near zone	Independent	+1.50	No	Yes
LentisMplus (LS-312MF 30, LS-313MF 30)	Oculentis GmbH	Ref, sector-shaped near zone	Independent	+3.00	No	Yes
LentisMplus toric (LS-312T1-T6, LS-313T1-T6)	Oculentis GmbH	Ref, sector-shaped near zone	Independent	+3.00	Yes	Yes
M-flex (580F, 630F)	Rayner Ltd.	Ref	Dependent	+3.00, +4.00	No	Yes
M-flex T (588F, 638F)	Rayner Ltd.	Ref	Dependent	+3.00, +4.00	Yes	Yes
MS 6125 Diff	Dr. Schmidt Intraocular Linsen	Diff	Dependent	+3.50	No	Yes
MS 614 Diff	Dr. Schmidt Intraocular Linsen	Diff, sulcus	Dependent	+3.50	No	Yes
MS 714 PB Diff	Dr. Schmidt Intraocular Linsen	Diff, sulcus, add-on	Dependent	+3.50	No	Yes
MS 714 TPB Diff	Dr. Schmidt Intraocular Linsen	Diff, sulcus, add-on	Dependent	+3.50	Yes	Yes
OptiVis	Aaren Scientific	Diff	Dependent	+2.80	No	Yes
PA 154N*	Allergan	Ref	Dependent	+3.50	No	No
PY-60MV*	Hoya	Ref	Dependent	+3.00	No	No
ReStor (SA60D3, SN60D3, MN60D3)	Alcon Laboratories	Diff + ref	Dependent	+4.00	No	No
ReStor (SN6AD1, SN6AD3)	Alcon Laboratories	Diff + ref	Dependent	+3.00, +4.00	No	Yes
ReStor (SND1-T2/3/4/5)	Alcon Laboratories	Diff + Ref	Dependent	+3.00	Yes	Yes
ReZoom (NXG1)	Abbott Medical Optics	Diff + ref	Dependent	+3.50	No	No
SFX MV1*	Hoya	Ref	Dependent	+2.25	No	No
Sulcoflex multifocal (653F)	Rayner Ltd.	Ref, sulcus, add-on	Dependent	+3.50	No	No
Sulcoflex multifocal toric (653Z)	Rayner Ltd.	Ref, sulcus, add-on	Dependent	+3.50	Yes	No
Tecnis (ZM900, ZMB00)	Abbott Medical Optics	Diff	Independent	+4.00	No	Yes
TrueVista 68STUV*	Storz	Ref	Dependent	+4.00	No	No

Diff = diffractive; IOL = intraocular lens; Ref = refractive

*No longer available

or edge in the material in which it travels scatters in numerous directions. Light energy arriving at an edge or discontinuity can thus be divided over 2 or more focal points, similar to refractive IOLs. Both effects were described by Fresnel in 1822¹⁸ when working on lenses for lighthouses and can be used to design IOLs with multiple focal points. The type of optics used influences the clinical results of the IOL, as will be described later.

More recently, so-called aspheric multifocal IOLs have been introduced. In these IOLs, optical properties of the IOL have been altered to decrease higher-order aberrations (HOAs) of the total optical system, primarily by compensating for the increased spherical aberration of the cornea in older subjects.^{19,20} Studies comparing aspheric and spherical monofocal IOLs

have reported superior visual performance of aspheric IOLs compared with their spherical counterparts, especially with respect to mesopic visual acuity and contrast sensitivity.^{21,22} In the case of multifocal IOLs, implantation of aspheric IOLs has been found to result in superior²³ or equal²⁴ visual performance compared with their spherical counterparts.

Apart from refractive versus diffractive designs and spherical versus aspheric designs, multifocal IOL designs can be described as pupil dependent or pupil independent. In zonal refractive designs and designs with a central diffractive structure, the division of the light energy is dependent on pupil size. Intraocular lens designs with a similar peripheral and central optical zone are pupil independent. The differences

between multifocal IOL designs are best illustrated in ray-tracing studies²⁵ and optomechanical eye-model studies.²⁶ This article describes the available designs as well as the results and side-effects of implanting multifocal IOLs following cataract surgery as reported in the English and German peer-reviewed literature.

MATERIALS AND METHODS

Bibliographic research was performed in Pubmed/Medline and most recently updated May 1, 2012. Keywords used were "multifocal intraocular lens," "multifocal intraocular lenses," "multifocal IOL," and the respective brand names shown in Table 1. Articles were included when they reported on clinical trials, adult patients with cataract, bilateral surgery with a single type of multifocal IOL, absence of coexisting ocular pathology such as amblyopia, and absence of previous or subsequent corneal refractive procedures such as limbal relaxing incisions or laser refractive surgery. Papers were classified as randomized controlled trials (RCTs) or nonrandomized case series, with or without a control group. The RCTs were included regardless of the date of publication; case series were included if published after January 1, 2009. Data analysis focused on uncorrected distance visual acuity (defined as visual acuity measured at 4 to 6 meters) and uncorrected near visual acuity (defined as 30 to 50 centimeters, standardized for all subjects in the study or at the working distance preferred by the individual patient). The mean visual acuities are reported as logMAR units \pm standard deviation, if necessary after conversion of reported alternative visual acuity units. A secondary outcome parameter noted was spectacle independence, defined as not using spectacles

for distance, intermediate, and near vision tasks. Photic phenomena, such as glare, flare, and halos, intermediate visual acuity, and contrast sensitivity were not used as outcome parameters given the lack of uniformity in reporting these findings in the available papers, as described later.

Bibliographic Research and Data Analysis

The search for RCTs reporting the results of multifocal IOL implantation after phacoemulsification of the crystalline lens identified 18 papers. One paper was excluded from the current review because it reported the results of multifocal IOL implantation in refractive lens exchange rather than in cataract surgery.²⁷ A second paper was excluded because it reported the results of unilateral multifocal IOL implantation.²⁸

The search for nonrandomized case series, with or without control groups, published after January 1, 2009, reporting the results of multifocal IOLs identified 128 papers. Eighty-seven were excluded for reasons presented in Table 2. The 41 remaining studies were included in the current literature review.

Visual Acuity and Spectacle Independence

The results of the papers reporting on RCTs are presented in Table 3.²⁹⁻⁴⁴ The results of the papers reporting on case series either comparing the results of different types of multifocal IOLs or reporting the results of a single type of multifocal IOL are presented in Table 4.^{23,45-85}

DISCUSSION

Multifocal IOL implantation is aimed at providing patients with good uncorrected visual acuity for both

Table 2. Reasons for exclusion of 87 studies from the literature review.

Reason for Exclusion	Number of Studies
Outcome parameters other than visual acuity and spectacle independence used (stereopsis, perimetry, electroretinography, effects of simulated astigmatism, pupillometry, contrast sensitivity, intraocular straylight, wavefront aberrometry)	16
In vitro study	13
Descriptions of postoperative complications (posterior capsule opacification, endophthalmitis, interference with intraoperative view during vitrectomy, autorefraction and optical coherence tomography, occurrence of dysphotopsia and other reasons for patient dissatisfaction)	11
Multifocal IOL combined with laser refractive surgery	8
Refractive lens exchange study population	5
Previously published results, comment on published paper, review	7
Unilateral and/or pediatric study population	6
Alternative techniques (scleral fixation of multifocal IOL, cyclosporine as adjuvant therapy)	4
Description of national practice patterns	2
Description of questionnaire for spectacle dependence	2
Monovision strategies	2
IOL of non-multifocal design	2
Different types of multifocal IOLs implanted in contralateral eyes	2
Cases of dissatisfied patients	2
Reporting visual acuities as median value only	1
Reporting visual acuities as percentages of patients only	1
Amblyopic cases	1
Age-related macular degeneration cases	1
Cost-benefit analysis	1

Table 3. Randomized controlled trials comparing results of implanting different types of multifocal IOLs.

Study*	Year	IOL Type (Number of Eyes)	UNVA (LogMar)	UDVA (LogMar)	Complete SI (% of Patients)
Santhiago ²⁸	2012	ReStor SN6AD1 (40)	0.022 ± 0.08	0.032 ± 0.07	90%
		ReStor SN6AD3 (40)	0.027 ± 0.02	0.023 ± 0.12	90%
Alio ²⁹	2011	ReStor SN6AD3 (38)	0.28 ± 0.04 logRAD	0.13 ± 0.13	-
		Acri.LISA 366D (42)	0.19 ± 0.08 logRAD	0.10 ± 0.11	-
Santhiago ³⁰	2011	ReStor SN6AD1 (20)	0.022 ± 0.08	0.032 ± 0.07	90%
		ReStor SN6AD3 (20)	0.027 ± 0.02	0.023 ± 0.12	90%
Alio ³¹	2011	ReStor SN6AD3 (78)	112 ± 22 wpm	0.15 ± 0.12	-
		Acri.LISA 366D (84)	115 ± 42 wpm	0.12 ± 0.11	-
		ReZoom NXG1 (70)	101 ± 16 wpm	0.12 ± 0.13	-
Alfonso ³²	2010	ReStor SN60D3 (20)	0.03 ± 0.05 [†]	-0.04 ± 0.10 [†]	-
		ReStor SN6AD3 (20)	-0.05 ± 0.06 [†]	0.08 ± 0.10 [†]	-
		ReStor SN6AD1 (20)	-0.08 ± 0.04 [†]	-0.06 ± 0.05 [†]	-
		Acri.LISA 366D (20)	-0.02 ± 0.08 [†]	-0.08 ± 0.08 [†]	-
Santhiago ³³	2010	ReStor SN6AD3 (32)	0.03 ± 0.08 [†]	0.02 ± 0.07 [†]	-
		ReStor SN6AD1 (32)	0.02 ± 0.08 [†]	0.03 ± 0.07 [†]	-
Maxwell ³⁴	2009	ReStor SN6AD3 (228)	0.12 [†]	0.02 [†]	81.2%
		ReStor SN6AD1 (232)	0.10 [†]	0.02 [†]	78.3%
Martínez-Palmer ³⁵	2008	Tecnis ZM900 (52)	0.06 ± 0.09 [‡]	0.18 ± 0.10	77.0%
		ReZoomNXG1 (64)	0.22 ± 0.14 [‡]	0.14 ± 0.12	44%
Cillino ³⁶	2008	Acri.Twin (64)	0.11 ± 0.12 [‡]	0.16 ± 0.12	87.5%
		Array SA40N (32)	0.20 ± 0.06	0.06 ± 0.10	43.7%
		ReZoomNXG1 (30)	0.21 ± 0.10	0.07 ± 0.14	53.3%
Hütz ³⁷	2008	Tecnis ZM900 (32)	0.14 ± 0.11	0.16 ± 0.10	87.5%
		Array SA40N (20)	0.43 ± 0.14	-	-
		ReStor SA60D3 (20)	0.28 ± 0.15	-	-
Gunenc ³⁸	2008	Array SA40N (20)	20% ≥ J1 [†]	90% ≥ 20/25	60%
			40% ≥ J2 [†]		
		CeeOn 811E (20)	90% ≥ J1 [†]	80% ≥ 20/25	60%
			100% ≥ J2 [†]		
Chiam ³⁹	2007	ReStor SA60D3 (100)	0.11	0.06	86%
		ReZoom NXG1 (100)	0.23	0.02	70%
Mester ⁴⁰	2007	Array SA40 (50)	0.40 [§]	0.08 ^{†,§}	33.3%
		Tecnis ZM900 (50)	0.22 [§]	0.08 ^{†,§}	82.6%
Hütz ⁴¹	2006	Array SA40N (20)	69 wpm [†]	—	—
		Tecnis ZM001 (20)	166 wpm [†]	—	—
		ReStor SA60D3 (20)	138 wpm [†]	—	—
Leyland ⁴²	2002	Array SA40NB (58)	0.43 ± 0.16	0.06 ± 0.10	28%
		TrueVista (30)	0.46 ± 0.21	0.10 ± 0.15	33%
Liekfeld ⁴³	1998	CeeOn 811E (26)	0.04 ± 0.05	0.09 ± 0.12	—
		PA 154N (24)	0.32 ± 0.24	0.12 ± 0.10	—

IOL = intraocular lens; J = Jaeger optotype; SI = spectacle independence; UDVA = uncorrected distance visual acuity; UNVA = uncorrected near visual acuity; wpm = words per minute

*First author

[†]Binocular

[‡]Binocular with distance correction

[§]Derived from figure

distance and near visual tasks. However, patients with a monofocal IOL can also have both good uncorrected distance and near visual acuity resulting from favorable corneal astigmatism^{86,87}; favorable corneal wavefront aberrations^{88,89}; or myopic undercorrection in 1 eye, resulting in pseudophakic monovision.^{50,90} The RCTs^{37,60,91-95} and metaanalyses of RCTs^{96,97}

comparing the results of multifocal IOL implantation with the results of monofocal IOL implantation conclude that uncorrected near vision is improved by implantation of a multifocal IOL, resulting in lower levels of spectacle dependence for near tasks without compromising distance visual acuity. The results of the current bibliographic search for papers in the peer-

Table 4. Nonrandomized studies reporting results of implanting different types of multifocal IOLs.

	Study*	Eyes	UNVA (LogMAR)	UDVA (LogMAR)	SI
Acri.LISA 366D	Alfonso ²²	40	-0.05 ± 0.07 [†]	0.01 ± 0.18 [†]	-
	Alió ⁴⁴	40	0.12 ± 0.12	0.10 ± 0.12	-
	Can ⁴⁵	30	0.08 ± 0.20	0.10 ± 0.07	100% (n) 96.6% (i) 100% (d)
	Castillo-Gómez ⁴⁶	20	0.06	0.15	-
	Fernández-Vega ⁴⁷	170	0.00 ± 0.02	0.07 ± 0.02	-
AcrivaReviol MFM 611	Can ⁴⁵	30	0.00 ± 0.03 0.02 ± 0.05	0.10 ± 0.16 0.07 ± 0.08	100% (n) 100% (i) 100% (d)
	Fujimoto ⁴⁸	72	0.24	0.06	34.7%
Array	Ito ⁴⁹	44	0.19 ± 0.12 [†]	-0.10 ± 0.00 [†]	-
	Mojzis ⁵⁰	23	0.24 ± 0.15	0.17 ± 0.13	-
AT LISA 909M		41	0.10 ± 0.09	0.12 ± 0.10	-
	Visser ⁵¹	45	0.20 ± 0.16	0.04 ± 0.15	53%
LentisMplus LS- 312 MF15	Alió ⁵²	22	0.45 ± 0.19	0.20 ± 0.14	-
	Alió ⁵³	43	0.21 ± 0.17	0.15 ± 0.21	-
LentisMplus LS- 312 MF30	Alió ⁵²	21	0.21 ± 0.10	0.14 ± 0.11	-
	Alió ⁵⁴	24	0.30 ± 0.21	0.25 ± 0.33	-
M-flex 630F	van der Linden ⁵⁵	90	0.16 ± 0.21	0.04 ± 0.15	-
	McAlinden ⁵⁶	44	162 wpm	0.04 ± 0.25	-
	Aslam ⁵⁷	20	65% ≥ J6	0.18 ± 0.20	-
	Cezón-Prieto ⁵⁸	32	0.28 ± 0.11	0.09 ± 0.09	70% (n) 80% (i) 90% (d)
	Gerten ⁵⁹	56	0.16 ± 0.13	0.10 ± 0.11	93.3%
MS 714 PB	Wolter-Roessler ⁶⁰	50	0.20	0.05	-
	Piovella ⁶¹	121	88.6% ≥ 0.10	84.1% ≥ 0.10	-
ReStor SA60D3, SN60D3, SN6AD3	Alfonso ²²	36	-0.04 ± 0.18 [†]	0.02 ± 0.13 [†]	-
	Alió ⁴⁴	40	0.19 ± 0.12	0.19 ± 0.18	-
	Blaylock ⁶²	74	0.06 [†]	0.00 [†]	-
	Chang ⁶³	30	0.07 [†]	0.08 [†]	72.7%
	Cionni ⁶⁴	190	0.11	0.05	80.6%
	Gierek-Ciacura ⁶⁵	20	0.11 ± 0.01	0.17 ± 0.02	80.0%
	Hayashi ⁶⁶	63	0.1 [§]	0.1 [§]	-
	Hida ⁶⁷	40	85% > J2	0.03 ± 0.05	-
	Mester ⁶⁸	40	0.24 ± 0.18 [‡]	0.17 ± 0.22	-
	Petermeier ⁶⁹	30	0.0 ± 0.07 [†]	0.0 ± 0.07 [†]	100% (n) 80% (i) 93% (d)
	de Vries ⁷⁰	46	0.01 ± 0.05 [†]	0.05 ± 0.12 [†]	-
	Zelichowska ⁷¹	46	-	0.03 ± 0.05	-

(continued on next page)

Table 4. (Cont.)

	Study*	Eyes	UNVA (LogMAR)	UDVA (LogMAR)	SI
ReStor	Alfonso ⁷²	40	-0.04 ± 0.06 [†]	0.00 ± 0.10 [†]	-
SN6AD1	Hayashi ⁷³	64	0.21	0.08	-
	Kohnen ⁷⁴	186	-0.01 ± 0.11 [†]	-0.03 ± 0.13 [†]	88%
	van der Linden ⁵⁵	143	0.05 ± 0.14	0.06 ± 0.25	-
	Mester ⁶⁸	40	0.17 ± 0.14 [‡]	0.14 ± 0.14 [‡]	-
	Petermeier ⁶⁹	24	0.1 ± 0.10 [†]	0.0 ± 0.05 [†]	92% (n) 83% (i) 92% (d)
	de Vries ⁷⁰	68	0.04 ± 0.12	0.04 ± 0.14	-
	Zhang ⁷⁵	42	100% ≥ 0.10	90% ≥ 0.10	81% (d)
	Chang ⁶³	30	0.17 [†]	-0.01 [†]	50%
	Forte ⁷⁶	55	12.3 ± 0.7	0.05 ± 0.09	-
	Gierek-Ciaciura ⁶⁵	20	0.20 ± 0.04	0.11 ± 0.01	70%
ReZoom NXG1	Lin ⁷⁷	28	153 ± 44 wpm	0.01 [§]	-
	Zelichowska ⁷¹	46	-	0.03 ± 0.06	-
	Hayashi ⁷⁸	44	0.38	0.08	-
SFX MV1	Khan ⁷⁹	4	100% ≥ J4	100% ≥ 0.10	-
Sulcoflex multifocal					
Tecnis ZM900, ZMB00	Akaishi ⁸⁰	2500	0.00 ± 0.00	0.06 ± 0.09	97.9 %
	Bautista ⁸¹	70	78.6% J1	0.076 ± 0.014	-
	Castillo-Gómez ⁴⁶	20	0.11	0.08	-
	Gierek-Ciaciura ⁶⁵	20	0.12 ± 0.03	0.14 ± 0.02	80 %
	Packer ⁸²	244	0.16	0.04	84.8%
	Palomina-Bautista ⁸³	250	0.22 ± 0.08	0.14 ± 0.10	88.4%
	Yoshino ⁸⁴	30	0.19 ± 0.10	0.01 ± 0.10	86.7%

d = distance; i = intermediate; IOL = intraocular lens; J = Jaeger optotype; n = near; RLE = refractive lens exchange; SI = spectacle independence; UDVA = uncorrected distance visual acuity; UNVA = uncorrected near visual acuity

*First author

[†]Binocular

[‡]Binocular with distance correction

[§]Derived from figure

reviewed literature reporting on the results of bilateral implantation of multifocal IOLs in cataract surgery demonstrate that implantation of both refractive^{37,64} and diffractive^{33,37,64,71,75,98} multifocal IOLs result in high levels of uncorrected distance and near visual acuity and therefore to increased levels of spectacle independence compared with monofocal IOLs.

Despite their benefits of uncorrected visual acuity at multiple distances, multifocal IOLs are associated with certain drawbacks. First, halos and glare are more often reported by patients with a multifocal IOL than with a monofocal IOL.^{99,100} Refractive multifocal IOLs appear to be associated with more photic phenomena than diffractive multifocal IOLs.³⁷ Photic

phenomena are among the most frequent reasons for dissatisfaction after multifocal IOL implantation.^{101,102}

Second, multifocal IOLs are associated with lower contrast sensitivity than monofocal IOLs.³⁷ Especially in mesopic circumstances^{74,103} and in patients with decreased contrast sensitivity due to ocular pathology such as macular degeneration or corneal dystrophies, this loss of contrast sensitivity can become clinically relevant.^{41,104,105} The reason for the lower contrast sensitivity could be that multifocal IOLs result in coexisting images, 1 sharp and 1 out of focus, with the light from the latter reducing the detectability of the former image. Diffractive multifocal IOLs appear to be equal or superior to refractive multifocal IOLs with respect

to contrast sensitivity.^{41,104,105} Although contrast sensitivity in individuals with multifocal IOLs is diminished compared with individuals with monofocal IOLs, it is generally within the normal range of contrast in age-matched phakic individuals.^{103,106} Multifocal IOLs, unlike accommodating IOLs, depend on 2 or more fixed focal points that each represent 2 fixed working distances (far and near) at which they deliver a sharp image to the retina (surrounded by a blurred retinal image or images resulting from the other focal point or points). Working distances between these "sweet spots" are associated with suboptimal visual acuity, potentially resulting in difficulties with computer work and similar activities.³⁸ Traditionally, refractive multifocal IOLs performed better at intermediate than near distance.^{24,107} For that reason, refractive multifocal IOLs were implanted bilaterally in patients with strong intermediate vision demands, were combined with a diffractive multifocal IOLs in the nondominant eye, or were combined with mini-monovision strategies leaving the nondominant eye slightly myopic to increase visual function at near distances.^{39,107,108} Similarly, diffractive multifocal IOLs have been combined with mini-monovision strategies to increase visual function at intermediate distances.⁶³ More recently, trifocal diffractive IOLs have been proposed to increase intermediate visual acuity.¹⁰⁹ The introduction of diffractive multifocal IOLs with lower near additions has increased visual acuity at intermediate distance without also decreasing near and distance visual acuity.^{33,69,75}

Comparing the performance of different types of multifocal IOLs is hampered for several reasons. First, despite work to develop instruments to measure subjective quality of vision,¹¹⁰ there is no consensus on which test or questionnaire to use for the measurement of the occurrence and severity of photic symptoms, resulting in the use of many different questionnaires and grading systems. Since photic phenomena such as glare and halos seem to wane with time,⁹⁸ a standardized follow-up time would also be essential to compare results of different IOLs in different studies. Second, there is no standardized test for near visual acuity. Some studies use single character reading charts such as Snellen³⁷ and Early Treatment of Diabetic Retinopathy Study⁶⁹ near visual acuity charts. Other studies use function-based tests such as the Minnesota Low Vision Reading Test chart,¹¹¹ the Radner chart,⁴² and variations thereof⁵⁰ measuring reading speed, number of mistakes, and critical character size when using sentences rather than single characters. Third, there is no consensus whether visual acuity should be measured binocularly or monocularly. Binocular visual acuity is generally higher, which might be the result of slight refractive differences between

the eyes (resulting in an effect comparable to pseudophakic monovision) or might be the result of more complex and less understood neurological processes. In clinical practice, binocular implantation has been shown to be preferable to monocular implantation.¹¹² Fourth, contrast sensitivity measurements are currently not standardized, with the CSV-1000,^{71,76} the Functional Acuity Test Chart,^{36,73,113} the Ginsburg box,^{41,69} and the CAT-2000^{74,79} systems being most widely used. Discussion exists of what levels should be considered normal given the large standard deviation of contrast sensitivity in normal subjects and whether multifocal IOLs should be compared with age-matched phakic subjects or age-matched subjects with a monofocal IOL. Finally, multifocal IOLs have been associated with higher levels of HOAs than monofocal IOLs.¹¹⁴ The role of these aberrations, however, is not clear. Not only has the value of HOAs for depth of focus been disputed,^{88,89} but the ability of wavefront analyzers to correctly measure aberrations in subjects with a multifocal IOL has not been clarified.^{115,116} Lower levels of HOAs in so-called aspheric optics have been shown to be beneficial in monofocal IOLs,²² but this is less evident in multifocal IOLs.^{24,117} Given the lack of consensus on any of these 5 items, a direct comparison of types of multifocal IOLs can be difficult. In a metaanalysis by Cochener et al.,¹¹⁸ a comparison of visual acuity, spectacle independence, and occurrence of halos with different multifocal IOL designs was performed using random effects Poisson regression models. Compared with refractive multifocal IOLs, diffractive multifocal IOLs were associated with a similar uncorrected distance visual acuity and superior near visual acuity resulting in higher spectacle independence. No significant differences were found in the incidence of halos with different types of multifocal IOLs.

In general, multifocal IOLs are able to provide patients with excellent uncorrected distance and near visual acuity resulting in high levels of spectacle independence. Although superior from a theoretical point of view, currently available accommodating IOLs are unable to offer the same level of near visual acuity.¹¹⁹ Dissatisfaction following implantation of multifocal IOLs is rare and is often amenable to treatment.^{101,102} Some cases of dissatisfaction are due to the occurrence of phenomena inherent to the design of multifocal IOLs (such as glare and halos) and are therefore more difficult to treat.^{101,102} This demonstrates the importance of preoperative patient education, careful selection of cases, and individualized weighing of benefits and side-effects of multifocal IOLs.^{94,120,121} If these principles are respected, multifocal IOLs can lead to excellent results and can be of great value to present-day ophthalmology.

REFERENCES

- Baumeister M, Kohnen T. Akkommodation und Presbyopie. Teil 1: Physiologie der Akkommodation und Entwicklung der Presbyopie [Accommodation and presbyopia. Part 1: physiology of accommodation and development of presbyopia]. *Ophthalmologie* 2008; 105:597–608; CME quiz 609–610
- Helmholtz H. Ueber die Akkommodation des Auges. *Albrecht von Graefes Arch Klin Exp Ophthalmol* 1855; 1(2):1–74
- Anderson HA, Glasser A, Manny RE, Stuebing KK. Age-related changes in accommodative dynamics from preschool to adulthood. *Invest Ophthalmol Vis Sci* 2010; 51:614–622. Available at: <http://www.iovs.org/content/51/1/614.full.pdf>. Accessed July 13, 2012
- Glasser A, Campbell MCW. Presbyopia and the optical changes in the human crystalline lens with age. *Vision Res* 1998; 38:209–229
- Glasser A, Campbell MCW. Biometric, optical and physical changes in the isolated human crystalline lens with age in relation to presbyopia. *Vision Res* 1999; 39:1991–2015
- Stachs O, Martin H, Behrend D, Schmitz KP, Guthoff R. Three-dimensional ultrasound biomicroscopy, environmental and conventional scanning electron microscopy investigations of the human zonula ciliaris for numerical modelling of accommodation. *Graefes Arch Clin Exp Ophthalmol* 2006; 244:836–844
- Stachs O, Martin H, Kirchhoff A, Stave J, Terwee T, Guthoff R. Monitoring accommodative ciliary muscle function using three-dimensional ultrasound. *Graefes Arch Clin Exp Ophthalmol* 2002; 240:906–912
- Sharma KK, Santhoshkumar P. Lens aging: effects of crystallins. *Biochim Biophys Acta* 2009; 1790:1095–1108. Available at: <http://www.ncbi.nlm.nih.gov/pmc/articles/PMC2743770/pdf/nihms119283.pdf>. Accessed July 13, 2012
- Nishi Y, Mireskandari K, Khaw P, Findl O. Lens refilling to restore accommodation. *J Cataract Refract Surg* 2009; 35:374–382
- Menapace R, Findl O, Kriechbaum K, Leydolt-Koepl C. Accommodating intraocular lenses: a critical review of present and future concepts. *Graefes Arch Clin Exp Ophthalmol* 2007; 245:473–489
- Marchini G, Pedrotti E, Modesti M, Visentin S, Tosi R. Anterior segment changes during accommodation in eyes with a monofocal intraocular lens: high-frequency ultrasound study. *J Cataract Refract Surg* 2008; 34:949–956
- Koepl C, Findl O, Kriechbaum K, Sacu C, Drexler W. Change in IOL position and capsular bag size with an angulated intraocular lens early after cataract surgery. *J Cataract Refract Surg* 2005; 31:348–353
- Koepl C, Findl O, Menapace R, Kriechbaum K, Wirtitsch M, Buehl W, Sacu S, Drexler W. Pilocarpine-induced shift of an accommodating intraocular lens: AT-45 Crystalens. *J Cataract Refract Surg* 2005; 31:1290–1297
- Findl O, Leydolt C. Meta-analysis of accommodating intraocular lenses. *J Cataract Refract Surg* 2007; 33:522–527
- Davison JA, Simpson MJ. History and development of the apodized diffractive intraocular lens. *J Cataract Refract Surg* 2006; 32:849–858
- Keates RH, Pearce JL, Schneider RT. Clinical results of the multifocal lens. *J Cataract Refract Surg* 1987; 13:557–560
- Hansen TE, Corydon L, Krag S, Thim K. New multifocal intraocular lens design. *J Cataract Refract Surg* 1990; 16:38–41
- Fresnel A. Mémoire Sur un Nouveau Système d'Éclairage des Phares; lu à l'Académie des Sciences le 29 juillet 1822. Paris, France, Imprimerie Royale, 1822; 7. Available at: <http://www.bibnum.education.fr/physique/optique/m%C3%A9moire-sur-un-nouveau-syst%C3%A8me-d%E2%80%99%C3%A9clairage-des-phares>. Accessed July 13, 2012
- Holladay JT, Piers PA, Koranyi G, van der Mooren M, Norrby NES. A new intraocular lens design to reduce spherical aberration of pseudophakic eyes. *J Refract Surg* 2002; 18:683–691
- Montés-Micó R, Ferrer-Blasco T, Cerviño A. Analysis of the possible benefits of aspheric intraocular lenses: review of the literature. *J Cataract Refract Surg* 2009; 35:172–181
- Kohnen T, Klapproth OK, Bühren J. Effect of intraocular lens asphericity on quality of vision after cataract removal; an intraindividual comparison. *Ophthalmology* 2009; 116:1697–1706
- Tzelikis PF, Akaishi L, Trindade FC, Boteon JE. Spherical aberration and contrast sensitivity in eyes implanted with aspheric and spherical intraocular lenses: a comparative study. *Am J Ophthalmol* 2008; 145:827–833
- Alfonso JF, Puchades C, Fernández-Vega L, Montés-Micó R, Valcárcel B, Ferrer-Blasco T. Visual acuity comparison of 2 models of bifocal aspheric intraocular lenses. *J Cataract Refract Surg* 2009; 35:672–676
- de Vries NE, Webers CAB, Verbakel F, de Brabander J, Berendschot TT, Cheng YYY, Doors M, Nuijts RMMA. Visual outcome and patient satisfaction after multifocal intraocular lens implantation: aspheric versus spherical design. *J Cataract Refract Surg* 2010; 36:1897–1904
- Terwee T, Weeber H, van der Mooren M, Piers P. Visualization of the retinal image in an eye model with spherical and aspheric, diffractive, and refractive multifocal intraocular lenses. *J Refract Surg* 2008; 24:223–232
- Gobbi PG, Fasce F, Bozza S, Calori G, Brancato R. Far and near visual acuity with multifocal intraocular lenses in an optomechanical eye model with imaging capability. *J Cataract Refract Surg* 2007; 33:1082–1094
- Barisić A, Dekaris I, Gabrić N, Bohac M, Romac I, Mravičić I, Lazić R. Comparison of diffractive and refractive multifocal intraocular lenses in presbyopia treatment. *Coll Antropol* 2008; 32(suppl 2):27–31. Available at: <http://hrcak.srce.hr/file/54732>. Accessed July 13, 2012
- Toto L, Falconio G, Vecchiarino L, Scordia V, Di Nicola M, Ballone E, Mastropasqua L. Visual performance and biocompatibility of 2 multifocal diffractive IOLs: six-month comparative study. *J Cataract Refract Surg* 2007; 33:1419–1425
- Santhiago MR, Wilson SE, Netto MV, Ghanem RC, Monteiro MLR, Bechara SJ, Espana EM, Mello GR, Kara-Junior N. Modulation transfer function and optical quality after bilateral implantation of a +3.00 D versus a +4.00 D multifocal intraocular lens. *J Cataract Refract Surg* 2012; 38:215–220
- Alió JL, Plaza-Puche AB, Piñero DP, Amparo F, Rodríguez-Prats JL, Ayala MJ. Quality of life evaluation after implantation of 2 multifocal intraocular lens models and a monofocal lens. *J Cataract Refract Surg* 2011; 37:638–648
- Santhiago MR, Wilson SE, Netto MV, Espíndola RF, Shah RA, Ghanem RC, Bechara SJ, Kara-Junior N. Visual performance of an apodized diffractive multifocal intraocular lens with a +3.00 addition: 1-year follow-up. *J Refract Surg* 2011; 27:899–906
- Alió JL, Grabner G, Plaza-Puche AB, Rasp M, Piñero DP, Seyeddain O, Rodríguez-Prats JL, Ayala MJ, Moreu R, Hohensinn M, Riha W, Dexl A. Postoperative bilateral reading performance with 4 intraocular lens models: six-month results. *J Cataract Refract Surg* 2011; 37:842–852
- Alfonso JF, Fernández-Vega L, Puchades C, Montés-Micó R. Intermediate visual function with different multifocal intraocular lens models. *J Cataract Refract Surg* 2010; 36:733–739

34. Santhiago MR, Netto MV, Espindola RF, Mazurek MG, Gomes B de AF, Parede TRR, Harooni H, Kara-Junior N. Comparison of reading performance after bilateral implantation of multifocal intraocular lenses with +3.00 or +4.00 diopter addition. *J Cataract Refract Surg* 2010; 36:1874–1879
35. Maxwell WA, Lane SS, Zhou F. Performance of presbyopia-correcting intraocular lenses in distance optical bench tests. *J Cataract Refract Surg* 2009; 35:166–171
36. Martínez Palmer A, Gómez Faiña P, España Albelda A, Comas Serrano M, Nahra Saad D, Castilla Céspedes M. Visual function with bilateral implantation of monofocal and multifocal intraocular lenses: a prospective, randomized, controlled clinical trial. *J Refract Surg* 2008; 24:257–264
37. Cillino S, Casuccio A, Di Pace F, Morreale R, Pillitteri F, Cillino G, Lodato G. One-year outcomes with new-generation multifocal intraocular lenses. *Ophthalmology* 2008; 115:1508–1516
38. Hütz WW, Eckhardt HB, Röhrig B, Grolmus R. Intermediate vision and reading speed with Array, Tecnis, and ReSTOR intraocular lenses. *J Refract Surg* 2008; 24:251–256
39. Gunenc U, Celik L. Long-term experience with mixing and matching refractive Array and diffractive CeeOn multifocal intraocular lenses. *J Refract Surg* 2008; 24:233–242
40. Chiam PJT, Chan JH, Haider SI, Karia N, Kasaby H, Aggarwal RK. Functional vision with bilateral ReZoom and ReSTOR intraocular lenses 6 months after cataract surgery. *J Cataract Refract Surg* 2007; 33:2057–2061
41. Mester U, Hunold W, Wesendahl T, Kaymak H. Functional outcomes after implantation of Tecnis ZM900 and Array SA40 multifocal intraocular lenses. *J Cataract Refract Surg* 2007; 33:1033–1040
42. Hütz WW, Eckhardt HB, Röhrig B, Grolmus R. Reading ability with 3 multifocal intraocular lens models. *J Cataract Refract Surg* 2006; 32:2015–2021
43. Leyland MD, Langan L, Goolfee F, Lee N, Bloom PA. Prospective randomised double-masked trial of bilateral multifocal, bifocal or monofocal intraocular lenses. *Eye* 2002; 16:481–490. Available at: <http://www.nature.com/eye/journal/v16/n4/pdf/6700077a.pdf>. Accessed July 13, 2012
44. Liekfeld A, Walkow T, Anders N, Pham DT, Wollensak J. Prospektiver Vergleich zweier Multifokallinsenmodelle [Prospective comparison of two multifocal lens models]. *Ophthalmologie* 1998; 95:253–256
45. Alió JL, Piñero DP, Plaza-Puche AB, Amparo F, Jiménez R, Rodríguez-Prats JL, Javaloy J. Visual and optical performance with two different diffractive multifocal intraocular lenses compared to a monofocal lens. *J Refract Surg* 2011; 27:570–581
46. Can İ, Bostancı Ceran B, Soyugelen G, Takmaz T. Comparison of clinical outcomes with 2 small-incision diffractive multifocal intraocular lenses. *J Cataract Refract Surg* 2012; 38:60–67
47. Castillo-Gómez A, Carmona-González D, Martínez-de-la-Casa JM, Palomino-Bautista C, García-Feijoo J. Evaluation of image quality after implantation of 2 diffractive multifocal intraocular lens models. *J Cataract Refract Surg* 2009; 35:1244–1250
48. Fernández-Vega L, Alfonso JF, Baamonde B, Madrid-Costa D, Montés-Micó R, Lozano J. Visual and refractive outcomes in hyperopic pseudophakic patients implanted with the Acri.LISA 366D multifocal intraocular lens. *Am J Ophthalmol* 2009; 148:214–220
49. Fujimoto K, Honda K, Wada YR, Tanaka M, Irie T. Four-year experience with a silicone refractive multifocal intraocular lens. *J Cataract Refract Surg* 2010; 36:1330–1335
50. Ito M, Shimizu K. Reading ability with pseudophakic monovision and with refractive multifocal intraocular lenses: comparative study. *J Cataract Refract Surg* 2009; 35:1501–1504
51. Mojzis P, Piñero DP, Studeny P, Tomás J, Korda V, Plaza AB, Alió JL. Comparative analysis of clinical outcomes obtained with a new diffractive multifocal toric intraocular lens implanted through two types of corneal incision. *J Refract Surg* 2011; 27:648–657
52. Visser N, Nuijts RMMA, de Vries NE, Bauer NJC. Visual outcomes and patient satisfaction after cataract surgery with toric multifocal intraocular lens implantation. *J Cataract Refract Surg* 2011; 37:2034–2042
53. Alió JL, Plaza-Puche AB, Piñero DP, Javaloy J, Ayala MJ. Comparative analysis of the clinical outcomes with 2 multifocal intraocular lens models with rotational asymmetry. *J Cataract Refract Surg* 2011; 37:1605–1614
54. Alió JL, Plaza-Puche AB, Piñero DP. Rotationally asymmetric multifocal IOL implantation with and without capsular tension ring: refractive and visual outcomes and intraocular optical performance. *J Refract Surg* 2012; 28:253–258
55. Alió JL, Piñero DP, Plaza-Puche AB, Rodríguez Chan MJ. Visual outcomes and optical performance of a monofocal intraocular lens and a new-generation multifocal intraocular lens. *J Cataract Refract Surg* 2011; 37:241–250
56. van der Linden JW, van Velthoven M, van der Meulen I, Nieuwendaal C, Mourits M, Lapid-Gortzak R. Comparison of a new-generation sectorial addition multifocal intraocular lens and a diffractive apodized multifocal intraocular lens. *J Cataract Refract Surg* 2012; 38:68–73
57. McAlinden C, Moore JE. Multifocal intraocular lens with a surface-embedded near section: short-term clinical outcomes. *J Cataract Refract Surg* 2011; 37:441–445
58. Aslam SA, Kashani S, Jones E, Claoué C. Pilot study and functional results following implantation of the M-flex 630F multifocal intraocular lens. *J Cataract Refract Surg* 2009; 35:792
59. Cezón Prieto J, Bautista MJ. Visual outcomes after implantation of a refractive multifocal intraocular lens with a +3.00 D addition. *J Cataract Refract Surg* 2010; 36:1508–1516
60. Gerten G, Kermani O, Schmiedt K, Farvili E, Foerster A, Oberheide U. Dual intraocular lens implantation: monofocal lens in the bag and additional diffractive multifocal lens in the sulcus. *J Cataract Refract Surg* 2009; 35:2136–2143
61. Wolter-Roessler M, Kuchle M. Implantation multifokaler Add-on-Intraokularlinsen simultan mit Katarakt-Operation: Ergebnisse einer prospektiven Studie [Implantation of multifocal add-on IOLs simultaneously with cataract surgery: results of a prospective study]. *Klin Monatsbl Augenheilkd* 2010; 227:653–656
62. Piovella M, Bosc J- M. Clinical evaluation of the OptiVis™ multifocal intraocular lens. *Adv Ther* 2011; 28:1012–1020
63. Blaylock JF, Si Z, Prescott C, Aitchison S. Intermediate optimization of vision with bilateral nonaspheric multifocal intraocular lens implantation. *J Cataract Refract Surg* 2009; 35:303–311
64. Chang DF. Prospective functional and clinical comparison of bilateral ReZoom and ReSTOR intraocular lenses in patients 70 years or younger. *J Cataract Refract Surg* 2008; 34:934–941
65. Cionni RJ, Chang DF, Donnenfeld ED, Lane SS, McCulley JP, Solomon KD. Clinical outcomes and functional visual performance: comparison of the ReSTOR apodized diffractive intraocular lens to a monofocal control. *Br J Ophthalmol* 2009; 93:1215–1219
66. Gierek-Ciaciura S, Cwalina L, Bednarski L, Mrukwa-Kominek E. A comparative clinical study of the visual results between three types of multifocal lenses. *Graefes Arch Clin Exp Ophthalmol* 2010; 248:133–140
67. Hayashi K, Masumoto M, Hayashi H. All-distance visual acuity in eyes with a nontinted or a yellow-tinted diffractive multifocal intraocular lens. *Jpn J Ophthalmol* 2009; 53:100–106

68. Hida WT, Motta AFP, Kara-José Junior N, Alves E, Tadeu M, Cordeiro LN, Nakano CT. Comparison between OPD-scan results and visual outcomes of monofocal and multifocal intraocular lenses. *Arq Bras Oftalmol* 2009; 72:526–532. Available at: <http://www.scielo.br/pdf/abo/v72n4/a17v72n4.pdf>. Accessed July 13, 2012
69. Mester U, Junker B, Kaymak H. Funktionelle Ergebnisse zweier Multifokallinsen mit unterschiedlicher Nahaddition [Functional results with two multifocal intraocular lenses with different near addition.]. *Ophthalmologe* 2011; 108:137–142
70. Petermeier K, Messias A, Gekeler F, Szurman P. Effect of +3.00 diopter and +4.00 diopter additions in multifocal intraocular lenses on defocus profiles, patient satisfaction, and contrast sensitivity. *J Cataract Refract Surg* 2011; 37:720–726
71. de Vries NE, Webers CAB, Montés-Micó R, Ferrer-Blasco T, Nuijts RMMA. Visual outcomes after cataract surgery with implantation of a +3.00 D or +4.00 D aspheric diffractive multifocal intraocular lens: comparative study. *J Cataract Refract Surg* 2010; 36:1316–1322
72. Żelichowska B, Rekas M, Stankiewicz A, Cerviño A, Montés-Micó R. Apodized diffractive versus refractive multifocal intraocular lenses: optical and visual evaluation. *J Cataract Refract Surg* 2008; 34:2036–2042
73. Alfonso JF, Fernández-Vega L, Amhaz H, Montés-Micó R, Valcárcel B, Ferrer-Blasco T. Visual function after implantation of an aspheric bifocal intraocular lens. *J Cataract Refract Surg* 2009; 35:885–892
74. Hayashi K, Manabe S-I, Hayashi H. Visual acuity from far to near and contrast sensitivity in eyes with a diffractive multifocal intraocular lens with a low addition power. *J Cataract Refract Surg* 2009; 35:2070–2076
75. Kohonen T, Nuijts R, Levy P, Haefliger E, Alfonso JF. Visual function after bilateral implantation of apodized diffractive aspheric multifocal intraocular lenses with a +3.0 D addition. *J Cataract Refract Surg* 2009; 35:2062–2069
76. Zhang F, Sugar A, Jacobsen G, Collins M. Visual function and patient satisfaction: comparison between bilateral diffractive multifocal intraocular lenses and monovision pseudophakia. *J Cataract Refract Surg* 2011; 37:446–453
77. Forte R, Ursoleo P. The ReZoom multifocal intraocular lens: 2-year follow-up. *Eur J Ophthalmol* 2009; 19:380–383
78. Lin H-T, Chen W-R, Ding Z-F, Chen W, Wu C-R. Clinical evaluation of two multifocal intraocular lens patterns. *Int J Ophthalmol* 2012; 5:76–83. Available at: <http://www.ncbi.nlm.nih.gov/pmc/articles/PMC3340833/pdf/ijo-05-01-076.pdf>. Accessed July 13, 2012
79. Hayashi K, Yoshida M, Hayashi H. All-distance visual acuity and contrast visual acuity in eyes with a refractive multifocal intraocular lens with minimal added power. *Ophthalmology* 2009; 116:401–408
80. Khan MI, Muhtaseb M. Performance of the Sulcoflex piggyback intraocular lens in pseudophakic patients. *J Refract Surg* 2011; 27:693–696
81. Akaishi L, Vaz R, Vilella G, Garcez RC, Tzelikis PF. Visual performance of Tecnis ZM900 diffractive multifocal IOL after 2500 implants: a 3-year followup. *J Ophthalmol* article ID 717591. Available at: <http://www.ncbi.nlm.nih.gov/pmc/articles/PMC2952904/pdf/JOP2010-717591.pdf>. Accessed July 13, 2012
82. Pallomino Bautista C, Carmona González D, Castillo Gómez A. Evolution of visual performance in 70 eyes implanted with the Tecnis[®] ZMB00 multifocal intraocular lens. *Clin Ophthalmol* 2012; 6:403–407. Available at: <http://www.dovepress.com/getfile.php?fileID=12348>. Accessed July 13, 2012
83. Packer M, Chu YR, Waltz KL, Donnenfeld ED, Wallace RB III, Featherstone K, Smith P, Bentow SS, Tarantino N. Evaluation of the aspheric Tecnis multifocal intraocular lens: one-year results from the first cohort of the Food and Drug Administration clinical trial. *Am J Ophthalmol* 2010; 149:577–584. Available at: <http://download.journals.elsevierhealth.com/pdfs/journals/0002-9394/PIIS0002939409008101.pdf>. Accessed July 13, 2012
84. Palomino Bautista C, Carmona González D, Castillo Gómez A, Christóbal Bescos JA. Evolution of visual performance in 250 eyes implanted with the Tecnis ZM900 multifocal IOL. *Eur J Ophthalmol* 2009; 19:762–768
85. Yoshino M, Bissen-Miyajima H, Oki S, Minami K, Nakamura K. Two-year follow-up after implantation of diffractive aspheric silicone multifocal intraocular lenses. *Acta Ophthalmol* 2011; 89:617–621
86. Verzella F, Calossi A. Multifocal effects of against-the-rule myopic astigmatism in pseudophakic eyes. *Refract Corneal Surg* 1993; 9:58–61
87. Nanavaty MA, Vasavada AR, Patel AS, Raj SM, Desai TH. Analysis of patients with good uncorrected distance and near vision after monofocal intraocular lens implantation. *J Cataract Refract Surg* 2006; 32:1091–1097
88. Rocha KM, Vabre L, Chateau N, Krueger RR. Expanding depth of focus by modifying higher-order aberrations induced by an adaptive optics visual simulator. *J Cataract Refract Surg* 2009; 35:1885–1892
89. Nanavaty MA, Spalton DJ, Boyce J, Saha S, Marshall J. Wavefront aberrations, depth of focus, and contrast sensitivity with aspheric and spherical intraocular lenses: fellow-eye study. *J Cataract Refract Surg* 2009; 35:663–671
90. Hayashi K, Yoshida M, Manabe S-I, Hayashi H. Optimal amount of anisometropia for pseudophakic monovision. *J Refract Surg* 2011; 27:332–338
91. Harman FE, Maling S, Kampougeris G, Langan L, Khan I, Lee N, Bloom PA. Comparing the 1CU accommodative, multifocal, and monofocal intraocular lenses; a randomized trial. *Ophthalmology* 2008; 115:993–1001
92. Zhao G, Zhang J, Zhou Y, Hu L, Che C, Jiang N. Visual function after monocular implantation of apodized diffractive multifocal or single-piece monofocal intraocular lens; randomized prospective comparison. *J Cataract Refract Surg* 2010; 36:282–285
93. Nijkamp MD, Dolders MGT, de Brabander J, van den Borne B, Hendrikse F, Nuijts RMMA. Effectiveness of multifocal intraocular lenses to correct presbyopia after cataract surgery; a randomized controlled trial. *Ophthalmology* 2004; 111:1832–1839
94. Sen HN, Sarikkola A-U, Uusitalo RJ, Laatikainen L. Quality of vision after AMO Array multifocal intraocular lens implantation. *J Cataract Refract Surg* 2004; 30:2483–2493
95. Zeng M, Liu Y, Liu X, Yuan Z, Luo L, Xia Y, Zeng Y. Aberration and contrast sensitivity comparison of aspherical and monofocal and multifocal intraocular lens eyes. *Clin Exp Ophthalmol* 2007; 35:355–360
96. Leyland M, Zinicola E. Multifocal versus monofocal intraocular lenses in cataract surgery; a systematic review. *Ophthalmology* 2003; 110:1789–1798
97. Leyland M, Pringle E. Multifocal versus monofocal intraocular lenses after cataract extraction. *Cochrane Database Syst Rev* 2006; (issue 4): art. no. CD003169
98. Webers CAB, Montés-Micó R, Tahzib NG, Cheng YYY, de Brabander J, Hendrikse F, Nuijts RMMA. Long-term follow-up of a multifocal apodized diffractive intraocular lens after cataract surgery. *J Cataract Refract Surg* 2008; 34:1476–1482
99. Chiam PJT, Chan JH, Aggarwal RK, Kasaby S. ReSTOR intraocular lens implantation in cataract surgery: quality of

- vision. *J Cataract Refract Surg* 2006; 32:1459–1463; errata, 1987
100. Häring G, Dick HB, Krummenauer F, Weissmantel U, Kröncke W. Subjective photic phenomena with refractive multifocal and monofocal intraocular lenses; results of a multicenter questionnaire. *J Cataract Refract Surg* 2001; 27:245–249
 101. Woodward MA, Randleman JB, Stulting RD. Dissatisfaction after multifocal intraocular lens implantation. *J Cataract Refract Surg* 2009; 35:992–997
 102. de Vries NE, Webers CAB, Touwslager WRH, Bauer NJC, de Brabander J, Berendschot TT, Nuijts RMMA. Dissatisfaction after implantation of multifocal intraocular lenses. *J Cataract Refract Surg* 2011; 37:859–865
 103. Alfonso JF, Fernández-Vega L, Merayo C, Montés-Micó R. Contrast sensitivity comparison between AcrySof ReSTOR and Acri.LISA aspheric intraocular lenses. *J Refract Surg* 2010; 26:471–477
 104. Mesci C, Erbil H, Özdöker L, Karakurt Y, Dolar Bilge A. Visual acuity and contrast sensitivity function after accommodative and multifocal intraocular lens implantation. *Eur J Ophthalmol* 2010; 20:90–100
 105. Mesci C, Erbil HH, Olgun A, Aydın N, Candemir B, Akçakaya AA. Differences in contrast sensitivity between monofocal, multifocal and accommodating intraocular lenses: long-term results. *Clin Exp Ophthalmol* 2010; 38:768–777
 106. Montés-Micó R, España E, Bueno I, Charman WN, Menezo JL. Visual performance with multifocal intraocular lenses: mesopic contrast sensitivity under distance and near conditions. *Ophthalmology* 2004; 111:85–96
 107. Pepose JS, Qazi MA, Davies J, Doane JF, Loden JC, Sivalingham V, Mahmoud AM. Visual performance of patients with bilateral vs combination Crystalens, ReZoom, and ReSTOR intraocular lens implants. *Am J Ophthalmol* 2007; 144:347–357
 108. Hütz WW, Bahner K, Röhrig B, Hengerer F. The combination of diffractive and refractive multifocal intraocular lenses to provide full visual function after cataract surgery. *Eur J Ophthalmol* 2010; 20:370–375
 109. Gatinel D, Pagnouille C, Houbrechts Y, Gobin L. Design and qualification of a diffractive trifocal optical profile for intraocular lenses. *J Cataract Refract Surg* 2011; 37:2060–2067
 110. McAlinden C, Pesudovs K, Moore JE. The development of an instrument to measure quality of vision: the Quality of Vision (QoV) questionnaire. *Invest Ophthalmol Vis Sci* 2010; 51:5537–5545. Available at: <http://www.iovs.org/content/51/11/5537.full.pdf>. Accessed July 13, 2012
 111. Brown D, Dougherty P, Gills JP, Hunkeler J, Sanders DR, Sanders ML. Functional reading acuity and performance: Comparison of 2 accommodating intraocular lenses. *J Cataract Refract Surg* 2009; 35:1711–1714
 112. Cionni RJ, Osher RH, Snyder ME, Nordlund ML. Visual outcome comparison of unilateral versus bilateral implantation of apodized diffractive multifocal intraocular lenses after cataract extraction: prospective 6-month study. *J Cataract Refract Surg* 2009; 35:1033–1039
 113. Alfonso JF, Fernández-Vega L, Señaris A, Montés-Micó R. Prospective study of the Acri.LISA bifocal intraocular lens. *J Cataract Refract Surg* 2007; 33:1930–1935
 114. Ortiz D, Alió JL, Bernabéu G, Pongo V. Optical performance of monofocal and multifocal intraocular lenses in the human eye. *J Cataract Refract Surg* 2008; 34:755–762
 115. Gatinel D. Limited accuracy of Hartmann-Shack wavefront sensing in eyes with diffractive multifocal IOLs [letter]. *J Cataract Refract Surg* 2008; 34:528; reply by L Toto, G Falconio, L Vecchiarino, V Scorcia, M Di Nicola, E Ballone, L Mastropasqua, 528–529
 116. Charman WN, Montés-Micó R, Radhakrishnan H. Problems in the measurement of wavefront aberration for eyes implanted with diffractive bifocal and multifocal intraocular lenses. *J Refract Surg* 2008; 24:280–286
 117. Felipe A, Pastor F, Artigas JM, Diez-Ajenjo A, Gené A, Menezo JL. Correlation between optics quality of multifocal intraocular lenses and visual acuity; tolerance to modulation transfer function decay. *J Cataract Refract Surg* 2010; 36:557–562
 118. Cochener B, Lafuma A, Khoshnood B, Courouve L, Berdeaux G. Comparison of outcomes with multifocal intraocular lenses: a meta-analysis. *Clin Ophthalmol* 2011; 5:45–56. Available at: <http://www.ncbi.nlm.nih.gov/pmc/articles/PMC3033003/pdf/oph-5-045.pdf>. Accessed July 13, 2012
 119. Patel S, Alió JL, Feinbaum C. Comparison of Acri. Smart multifocal IOL, Crystalens AT-45 accommodative IOL, and Technovision presbyLASIK for correcting presbyopia. *J Refract Surg* 2008; 24:294–299
 120. Pepose JS. Maximizing satisfaction with presbyopia-correcting intraocular lenses: the missing links. *Am J Ophthalmol* 2008; 146:641–648
 121. Lichtinger A, Rootman DS. Intraocular lenses for presbyopia correction: past, present, and future. *Curr Opin Ophthalmol* 2012; 23:40–46



First author:
Niels E. de Vries, MD, PhD, FEBO

*Department of Ophthalmology,
University Hospital Maastricht,
the Netherlands*